

# CERTIFICATE OF CALIBRATION

**Issued By** PASS Ltd - www.calibrate.co.uk - 0845 365 39 44

**Date of Issue** 18 September 2013

Certificate Number  
STD31072

Page 1 of 2 Pages



1 Alberto Street  
Stockton On Tees  
Teesside

TS18 2BQ

Approved Signatory

P.Beswick     I.Ibrahim     D.Kendrew

**Customer :** Body Comfort s.r.o.  
Maria Karpatova  
1 Page Hill, Ware  
Hertfordshire SG12 0RZ  
**Date Received :** 05 September 2013

<b>Instrument -</b>	System ID :	ID25524	Job Number :	J35696-1
	Description :	IR thermometer		
	Manufacturer :	Body Comfort		
	Model Number :	BC 07		
	Serial Number :	130325191		
	Procedure Version :	1.18		

## Environmental Conditions

Temperature : 20°C +/- 1°C  
Relative Humidity : 50% +/- 10%

Mains Voltage : 240V +/- 10V  
Mains Frequency : 50Hz +/- 1Hz

## Comments

Instrument was placed in lab and allowed to stabilise before calibration.

## Traceability Information

<i>Instrument description</i>	<i>Serial number</i>	<i>Certificate number</i>	<i>Cal. Date</i>	<i>Cal. Period</i>
ETI Black Body Calibrator	10118975	261011101	09/10/2012	52
ETI Reference Thermometer	D11370422	051012/N6	05/10/2012	52

**Calibrated By :** P.Beswick

**Date of Calibration :** 18 September 2013

This certificate provides traceability of measurement to recognised National Standards, and to the units of measurement realised at the National Physical Laboratory or other recognised National Standards laboratories.

Copyright of this certificate is owned by the issuing laboratory and may not be reproduced except with the prior written approval of the issuing laboratory. This certificate complies with the requirements of BS EN ISO 10012:2003.

# CERTIFICATE OF CALIBRATION

Certificate Number  
STD31072

Page 2 of 2 Pages

Test Title	Tolerance	Applied Value	Reading	Pass/Fail
<b>Temperature</b>				
0°C	300m°C	0.00°C	0.2°C	Pass
10°C	300m°C	10.00°C	9.8°C	Pass
20°C	300m°C	20.00°C	19.9°C	Pass
30°C	300m°C	30.00°C	30.1°C	Pass
36°C	300m°C	36.00°C	36.2°C	Pass
38°C	300m°C	38.00°C	38.1°C	Pass
39°C	300m°C	39.00°C	39.1°C	Pass
40°C	300m°C	40.00°C	39.8°C	Pass
50°C	300m°C	50.00°C	49.9°C	Pass
60°C	300m°C	60.00°C	60.2°C	Pass

**End Of Tests**

---

**Uncertainties**

---